



### Product Code

|               |                                        |
|---------------|----------------------------------------|
| UTC-0518      | Sieve Segregation Test Set             |
| UTG-4PC0050/Y | Sieve Ø 300x75 mm 5 mm Square Aperture |
| UTG-4002/Y    | Pan Ø 300x75 mm                        |


### Standards

EN 12350-11

The UTC-0518 Sieve Segregation Test Set is used for determining the resistance to sieve segregation of Self Compacting Concrete.

The Sieve has 5 mm square apertures with a frame diameter of 300 mm conforming to ISO 3310-2 standard.

Supplied complete with a Pan.



|                  |                |
|------------------|----------------|
| Dimensions       | 320x320x300 mm |
| Weight (approx.) | 2 kg           |